

## Datasheet for ABIN5654956 **HSPBP1 ELISA Kit**



Overview

Quantity:	96 tests
Target:	HSPBP1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

## Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Heat Shock 70 kDa Binding Protein 1 (HSPBP1). No significant cross-reactivity or interference between Heat Shock 70 kDa Binding Protein 1 (HSPBP1) and analogues was observed.
Sensitivity:	0.118 ng/mL
Target Details	

Target:	HSPBP1
Alternative Name:	Heat Shock 70kDa Binding Protein 1 (HSPBP1 Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5654956 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details	
Background:	Gene Name: Heat Shock 70 kDa Binding Protein 1
	Gene Aliases: FES1, Hsp70 Interacting Protein, Hsp70 Binding Protein,Cytoplasmic
	Cochaperone 1, Heat shock protein-binding protein 1, Hsp70-interacting protein 1
Gene ID:	23640
UniProt:	Q9NZL4
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Heat Shock 70 kDa Binding Protein 1 (HSPBP1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Heat Shock 70 kDa Binding Protein 1 (HSPBP1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Heat Shock 70 kDa Binding Protein 1 (HSPBP1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Heat Shock 70 kDa Binding Protein 1 (HSPBP1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Heat Shock 70kDa Binding Protein 1 (HSPBP1) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Heat Shock 70kDa Binding Protein 1 (HSPBP1) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5654956 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Application Details	
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months