

Datasheet for ABIN5655083 HGFA ELISA Kit



Overview

Quantity:	96 tests
Target:	HGFA
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	15.62 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.62 pg/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum, Tissue Homogenate	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of Hepatocyte Growth Factor Activator (HGFAC). No significant cross-reactivity or interference between Hepatocyte Growth Factor Activator (HGFAC) and analogues was observed.	
Sensitivity:	6.7 pg/mL	
Target Details		
Target:	HGFA	
Alternative Name:	Hepatocyte Growth Factor Activator (HGFA Products)	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5655083 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details

Background:

Gene Name: Hepatocyte Growth Factor Activator Gene Aliases: HGFA, HGFAP, HGF Activator

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than		
	5 % within the expiration date under appropriate storage condition. To minimize extra influence		
	on the performance, operation procedures and lab conditions, especially room temperature, air		
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that		
	the whole assay is performed by the same operator from the beginning to the end.		
Assay Time:	3 h		
Plate:	Pre-coated		
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate		
	provided in this kit has been pre-coated with an antibody specific to Hepatocyte Growth Factor		
	Activator (HGFAC). Standards or samples are then added to the appropriate microtiter plate		
	wells with a biotin-conjugated antibody specific to Hepatocyte Growth Factor Activator		
	(HGFAC). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate		
	well and incubated. After TMB substrate solution is added, only those wells that contain		
	Hepatocyte Growth Factor Activator (HGFAC), biotin-conjugated antibody and enzyme-		
	conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by		
	the addition of sulphuric acid solution and the color change is measured		
	spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of Hepatocyte		
	Growth Factor Activator (HGFAC) in the samples is then determined by comparing the O.D. of		
	the samples to the standard curve.		
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level		
	Hepatocyte Growth Factor Activator (HGFAC) were tested 20 times on one plate, respectively		
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level		
	Hepatocyte Growth Factor Activator (HGFAC) were tested on 3 different plates, 8 replicates in		
	each plate. CV(%) = SD/meanX100		
	Intra-Assay: CV<10%		
	Inter-Assay: CV<12%		
Restrictions:	For Research Use only		

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5655083 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Н	land	lina
11	ana	mig

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months