

## Datasheet for ABIN5655084 HGFA ELISA Kit



Overview

Quantity:	96 tests
Target:	HGFA
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	15.62 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.62 pg/mL
Application:	ELISA

## Product Details

Sample Type:	Plasma, Serum, Tissue Homogenate	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of Hepatocyte Growth Factor Activator (HGFAC). No significant cross-reactivity or interference between Hepatocyte Growth Factor Activator (HGFAC) and analogues was observed.	
Sensitivity:	5.7 pg/mL	
Target Details		
Target:	HGFA	
Alternative Name:	Hepatocyte Growth Factor Activator (HGFA Products)	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5655084 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details		
Background:	Gene Name: Hepatocyte Growth Factor Activator	
	Gene Aliases: HGFA, HGFAP, HGF Activator	
Gene ID:	58947	
UniProt:	Q5EBA7	
Application Details		
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than	
	5 % within the expiration date under appropriate storage condition. To minimize extra influence	
	on the performance, operation procedures and lab conditions, especially room temperature, air	
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that	
	the whole assay is performed by the same operator from the beginning to the end.	
Assay Time:	3 h	
Plate:	Pre-coated	
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate	
	provided in this kit has been pre-coated with an antibody specific to Hepatocyte Growth Factor	
	Activator (HGFAC). Standards or samples are then added to the appropriate microtiter plate	
	wells with a biotin-conjugated antibody specific to Hepatocyte Growth Factor Activator	
	(HGFAC). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate	
	well and incubated. After TMB substrate solution is added, only those wells that contain	
	Hepatocyte Growth Factor Activator (HGFAC), biotin-conjugated antibody and enzyme-	
	conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by	
	the addition of sulphuric acid solution and the color change is measured	
	spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of Hepatocyte	
	Growth Factor Activator (HGFAC) in the samples is then determined by comparing the O.D. of	
	the samples to the standard curve.	
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level	
	Hepatocyte Growth Factor Activator (HGFAC) were tested 20 times on one plate, respectively	
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level	
	Hepatocyte Growth Factor Activator (HGFAC) were tested on 3 different plates, 8 replicates in	
	each plate. CV(%) = SD/meanX100	
	Intra-Assay: CV<10%	
	Inter-Assay: CV<12%	
Restrictions:	For Research Use only	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5655084 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Н	land	lina
11	ana	mig

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months