

Datasheet for ABIN5655197 **Homovanillic Acid ELISA Kit**



[Go to Product page](#)

Overview

| | |
|--------------------------|-------------------------|
| Quantity: | 96 tests |
| Target: | Homovanillic Acid (HVA) |
| Reactivity: | Various Species |
| Method Type: | Competition ELISA |
| Detection Range: | 2.47 ng/mL - 200 ng/mL |
| Minimum Detection Limit: | 2.47 ng/mL |
| Application: | ELISA |

Product Details

| | |
|--------------------|---|
| Sample Type: | Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate |
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Specificity: | This assay has high sensitivity and excellent specificity for detection of Homovanillic Acid (HVA). No significant cross-reactivity or interference between Homovanillic Acid (HVA) and analogues was observed. |
| Sensitivity: | 0.97 ng/mL |

Target Details

| | |
|-------------------|--|
| Target: | Homovanillic Acid (HVA) |
| Alternative Name: | Homovanillic Acid (HVA Products) |

Target Details

| | |
|--------------|--|
| Target Type: | Chemical |
| Background: | Gene Name: Homovanillic Acid Gene Aliases: HOC6H3(OCH3)CH2COOH, 3-Methoxy-4-hydroxyphenyl Acetic Acid, 4-Hydroxy-3-Methoxy-Benzeneacetic Acid, 4-Hydroxy-3-Methoxyphenylacetic Acid |

Application Details

| | |
|------------------|---|
| Comment: | The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end. |
| Assay Time: | 2 h |
| Plate: | Pre-coated |
| Protocol: | This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Homovanillic Acid (HVA) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Homovanillic Acid (HVA) and unlabeled Homovanillic Acid (HVA) (Standards or samples) with the pre-coated antibody specific to Homovanillic Acid (HVA). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Homovanillic Acid (HVA) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Homovanillic Acid (HVA) in the sample. |
| Assay Precision: | Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Homovanillic Acid (HVA) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Homovanillic Acid (HVA) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = \frac{SD}{mean} \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$ |
| Restrictions: | For Research Use only |

Handling

| | |
|------------------|--|
| Handling Advice: | The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption. |
| Storage: | 4 °C,-20 °C |
| Storage Comment: | -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles. |
| Expiry Date: | 4-8 months |