



[Go to Product page](#)

Datasheet for ABIN5655328 Indole 3 Acetic Acid ELISA Kit

Overview

Quantity:	96 tests
Target:	Indole 3 Acetic Acid (IAA)
Reactivity:	Various Species
Method Type:	Competition ELISA
Detection Range:	2.47 ng/mL - 200 ng/mL
Minimum Detection Limit:	2.47 ng/mL
Application:	ELISA

Product Details

Sample Type:	Cell Culture Supernatant, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Indole 3 Acetic Acid (IAA). No significant cross-reactivity or interference between Indole 3 Acetic Acid (IAA) and analogues was observed.
Sensitivity:	0.89 ng/mL

Target Details

Target:	Indole 3 Acetic Acid (IAA)
Alternative Name:	Indole 3 Acetic Acid (IAA Products)

Target Details

Target Type:	Chemical
Background:	Gene Name: Indole 3 Acetic Acid Gene Aliases: IA-A, Indolyacetic Acid, Indoleacetic Acid, Heteroauxin

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 h
Plate:	Pre-coated
Protocol:	This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Indole 3 Acetic Acid (IAA) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Indole 3 Acetic Acid (IAA) and unlabeled Indole 3 Acetic Acid (IAA) (Standards or samples) with the pre-coated antibody specific to Indole 3 Acetic Acid (IAA). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Indole 3 Acetic Acid (IAA) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Indole 3 Acetic Acid (IAA) in the sample.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Indole 3 Acetic Acid (IAA) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Indole 3 Acetic Acid (IAA) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = \frac{SD}{mean} \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Restrictions:	For Research Use only

Handling

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and
------------------	--

Handling

specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Storage: 4 °C, -20 °C

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months