

## Datasheet for ABIN5655572

# **IFNAR2 ELISA Kit**



#### Overview

Quantity:	96 tests
Target:	IFNAR2
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.625 ng/mL - 40 ng/mL
Minimum Detection Limit:	0.625 ng/mL
Application:	ELISA

### **Product Details**

Sample Type:	Plasma, Serum, Urine
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Interferon Alpha/Beta Receptor 2 (IFNa/bR2). No significant cross-reactivity or interference between Interferon Alpha/Beta Receptor 2 (IFNa/bR2) and analogues was observed.
Sensitivity:	0.262 ng/mL

## Target Details

Target:	IFNAR2
Alternative Name:	Interferon Alpha/Beta Receptor 2 (IFNAR2 Products)

## Target Details

Background:	Gene Name: Interferon Alpha/Beta Receptor 2
	Gene Aliases: IFNAR2, IFN-alpha-REC, IFN-R, IFNABR, IFNARB, Interferon(Alpha,Beta And
	Omega)Receptor Beta Chain, Interferon alpha binding protein, Type I interferon receptor 2
Gene ID:	3455
UniProt:	P48551
Pathways:	JAK-STAT Signaling, Hepatitis C
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5% within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Interferon Alpha/Beta
	Receptor 2 (IFNa/bR2). Standards or samples are then added to the appropriate microtiter plate
	wells with a biotin-conjugated antibody specific to Interferon Alpha/Beta Receptor 2
	(IFNa/bR2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each
	microplate well and incubated. After TMB substrate solution is added, only those wells that
	contain Interferon Alpha/Beta Receptor 2 (IFNa/bR2), biotin-conjugated antibody and enzyme-
	conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by
	the addition of sulphuric acid solution and the color change is measured
	spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of Interferon
	Alpha/Beta Receptor 2 (IFNa/bR2) in the samples is then determined by comparing the O.D. of
	the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Interferon Alpha/Beta Receptor 2 (IFNa/bR2) were tested 20 times on one plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Interferon Alpha/Beta Receptor 2 (IFNa/bR2) were tested on 3 different plates, 8 replicates in
	each plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%

## **Application Details**

	Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months