

Datasheet for ABIN5655931 IL28A CLIA Kit



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	IL28A
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.41 pg/mL - 300 pg/mL
Minimum Detection Limit:	0.41 pg/mL
Application:	ELISA

Product Details

Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Specificity:	This assay has high sensitivity and excellent specificity for detection of Interleukin 28A (IL28A). No significant cross-reactivity or interference between Interleukin 28A (IL28A) and analogues was observed.
Sensitivity:	0.17 pg/mL

Target Details

Target:	IL28A
Alternative Name:	Interleukin 28A (IL28A Products)

Target Details

Background:	Gene Name: Interleukin 28A Gene Aliases: IFNL2, Interferon,Lambda 2, Cytokine Zcyto20
Gene ID:	282616
UniProt:	Q8IZJ0

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 - 3 h
Plate:	Pre-coated
Protocol:	The microplate provided in this kit has been pre-coated with an antibody specific to Interleukin 28A (IL28A). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Interleukin 28A (IL28A). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Interleukin 28A (IL28A) level in the sample or standard.,
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Interleukin 28A (IL28A) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Interleukin 28A (IL28A) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only

Handling

Handling Advice:	Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed
------------------	---

Handling

without interruption.

Storage: 4 °C,-20 °C

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months