

Datasheet for ABIN5655951

IL32 CLIA Kit



Overview

| Quantity: | 96 tests |
|--------------------------|--------------------------|
| Target: | IL32 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 15.62 pg/mL - 1000 pg/mL |
| Minimum Detection Limit: | 15.62 pg/mL |
| Application: | ELISA |

Product Details

| Sample Type: | Plasma, Serum |
|--------------------|---|
| Analytical Method: | Quantitative |
| Detection Method: | Chemiluminescent |
| Specificity: | This assay has high sensitivity and excellent specificity for detection of Interleukin 32 (IL32). No significant cross-reactivity or interference between Interleukin 32 (IL32) and analogues was observed. |
| Sensitivity: | 0.5 pg/mL |

Target Details

| Target: | IL32 |
|-------------------|--------------------------------|
| Alternative Name: | Interleukin 32 (IL32 Products) |

Target Details

| Background |
|------------|
|------------|

Gene Name: Interleukin 32

Gene Aliases: NK4, TAIF, TAIFb, TAIFd, Natural Killer Cell Transcript 4, Natural killer cells protein

4, Tumor necrosis factor alpha-inducing factor

Application Details

Comment:

The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time:

2 - 3 h

Plate:

Pre-coated

Protocol:

The microplate provided in this kit has been pre-coated with an antibody specific to Interleukin 32 (IL32). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Interleukin 32 (IL32). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Interleukin 32 (IL32) level in the sample or standard.,

Assay Precision:

Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Interleukin 32 (IL32) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Interleukin 32 (IL32) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%

Restrictions:

For Research Use only

Handling

Handling Advice:

Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Storage:

4 °C,-20 °C

Handling

| Expiry Date: | 4-8 months |
|------------------|---|
| | pack. Minimize freeze/thaw cycles. |
| | 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant |
| Storage Comment: | -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at |