

Datasheet for ABIN5656148 **Kallikrein 12 ELISA Kit**



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	Kallikrein 12 (KLK12)
Reactivity:	Human
Method Type:	Competition ELISA
Detection Range:	12.35 ng/mL - 1000 ng/mL
Minimum Detection Limit:	12.35 ng/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Kallikrein 12 (KLK12). No significant cross-reactivity or interference between Kallikrein 12 (KLK12) and analogues was observed.
Sensitivity:	7.14 ng/mL

Target Details

Target:	Kallikrein 12 (KLK12)
Alternative Name:	Kallikrein 12 (KLK12 Products)

Target Details

Background:	Gene Name: Kallikrein 12 Gene Aliases: KLK-L5, KLKL5, Kallikrein-Related Peptidase 12, Kallikrein-like protein 5
Gene ID:	43849
UniProt:	Q9UKR0
Pathways:	Complement System

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 h
Plate:	Pre-coated
Protocol:	This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Kallikrein 12 (KLK12) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Kallikrein 12 (KLK12) and unlabeled Kallikrein 12 (KLK12) (Standards or samples) with the pre-coated antibody specific to Kallikrein 12 (KLK12). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Kallikrein 12 (KLK12) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Kallikrein 12 (KLK12) in the sample.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Kallikrein 12 (KLK12) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Kallikrein 12 (KLK12) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = \frac{SD}{mean} \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Restrictions:	For Research Use only

Handling

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months