

Datasheet for ABIN5656421 LIM Domain Binding 1 Protein ELISA Kit



Overview

Quantity:	96 tests
Target:	LIM Domain Binding 1 Protein (LDB1)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA
Product Details	
Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of LIM Domain Binding

Protein 1 (LDB1). No significant cross-reactivity or interference between LIM Domain Binding
Protein 1 (LDB1) and analogues was observed.

0.061 ng/mL

Target Details

Sensitivity:

Target:	LIM Domain Binding 1 Protein (LDB1)
Alternative Name:	LIM Domain Binding Protein 1 (LDB1 Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5656421 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details	
Background:	Gene Name: LIM Domain Binding Protein 1 Gene Aliases: CLIM2, NLI, Carboxy Terminal LIM Domain Protein 2, LIM domain-binding factor CLIM2, Nuclear LIM interactor
Gene ID:	8861
UniProt:	Q86U70
Pathways:	Stem Cell Maintenance, Chromatin Binding

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to LIM Domain Binding
	Protein 1 (LDB1). Standards or samples are then added to the appropriate microtiter plate wells
	with a biotin-conjugated antibody specific to LIM Domain Binding Protein 1 (LDB1). Next, Avidin
	conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated.
	After TMB substrate solution is added, only those wells that contain LIM Domain Binding
	Protein 1 (LDB1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a
	change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid
	solution and the color change is measured spectrophotometrically at a wavelength of 450nm \pm
	10nm. The concentration of LIM Domain Binding Protein 1 (LDB1) in the samples is then
	determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level LIM
	Domain Binding Protein 1 (LDB1) were tested 20 times on one plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	LIM Domain Binding Protein 1 (LDB1) were tested on 3 different plates, 8 replicates in each
	plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%
	Inter-Assay: CV<12%

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5656421 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Application Details	
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months