

Datasheet for ABIN5656917

MTFMT ELISA Kit



_				
	۱۱ / ۱	rv		۱۸/
	' V '	 ı v	Ι.	v v

Quantity:	96 tests
Target:	MTFMT
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

Product Details

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Mitochondrial Methionyl tRNA Formyltransferase (MTFMT). No significant cross-reactivity or interference between Mitochondrial Methionyl tRNA Formyltransferase (MTFMT) and analogues was observed.
Sensitivity:	0.055 ng/mL

Target Details

Target:	MTFMT
Alternative Name:	Mitochondrial Methionyl tRNA Formyltransferase (MTFMT Products)

Target Details

l arget Details	
Background:	Gene Name: Mitochondrial Methionyl tRNA Formyltransferase
	Gene Aliases: FMT1
Gene ID:	69606
UniProt:	Q9D799
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5% within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Mitochondrial Methionyl
	tRNA Formyltransferase (MTFMT). Standards or samples are then added to the appropriate
	microtiter plate wells with a biotin-conjugated antibody specific to Mitochondrial Methionyl
	tRNA Formyltransferase (MTFMT). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is
	added to each microplate well and incubated. After TMB substrate solution is added, only those
	wells that contain Mitochondrial Methionyl tRNA Formyltransferase (MTFMT), biotin-
	conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-
	substrate reaction is terminated by the addition of sulphuric acid solution and the color change
	is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of
	Mitochondrial Methionyl tRNA Formyltransferase (MTFMT) in the samples is then determined
	by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Mitochondrial Methionyl tRNA Formyltransferase (MTFMT) were tested 20 times on one plate,
	respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Mitochondrial Methionyl tRNA Formyltransferase (MTFMT) were tested on 3 different plates, 8
	replicates in each plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%
	Inter-Assay: CV<12%

Application Details

Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and
	specimen samples should be assayed in duplicate. Once the procedure has been started, all
	steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at
	4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant
	pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months