

## Datasheet for ABIN5657089 MYH2 ELISA Kit



Go to Product page

Overview

Quantity:	96 tests
Target:	MYH2
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

## Product Details

Sample Type:	Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Myosin Heavy Chain 2, Skeletal Muscle, Adult (MYH2). No significant cross-reactivity or interference between Myosin Heavy Chain 2, Skeletal Muscle, Adult (MYH2) and analogues was observed.
Sensitivity:	0.117 ng/mL
Target Details	
Target:	MYH2

Alternative Name:	Myosin Heavy Chain 2, Skeletal Muscle, Adult (MYH2 Products)
-------------------	--

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5657089 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

## Target Details

Background:

Gene Name: Myosin Heavy Chain 2, Skeletal Muscle, Adult Gene Aliases: MYH2A, MYHSA2, MYHas8, MyHC-2A, MyHC-IIa

## Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than	
comment.	5 % within the expiration date under appropriate storage condition. To minimize extra influence	
	on the performance, operation procedures and lab conditions, especially room temperature, air	
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that	
	the whole assay is performed by the same operator from the beginning to the end.	
Assay Time:	3 h	
Plate:	Pre-coated	
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate	
	provided in this kit has been pre-coated with an antibody specific to Myosin Heavy Chain 2,	
	Skeletal Muscle, Adult (MYH2). Standards or samples are then added to the appropriate	
	microtiter plate wells with a biotin-conjugated antibody specific to Myosin Heavy Chain 2,	
	Skeletal Muscle, Adult (MYH2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is	
	added to each microplate well and incubated. After TMB substrate solution is added, only those	
	wells that contain Myosin Heavy Chain 2, Skeletal Muscle, Adult (MYH2), biotin-conjugated	
	antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate	
	reaction is terminated by the addition of sulphuric acid solution and the color change is	
	measured spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of	
	Myosin Heavy Chain 2, Skeletal Muscle, Adult (MYH2) in the samples is then determined by	
	comparing the O.D. of the samples to the standard curve.	
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level	
	Myosin Heavy Chain 2, Skeletal Muscle, Adult (MYH2) were tested 20 times on one plate,	
	respectively	
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level	
	Myosin Heavy Chain 2, Skeletal Muscle, Adult (MYH2) were tested on 3 different plates, 8	
	replicates in each plate. CV(%) = SD/meanX100	
	Intra-Assay: CV<10%	
	Inter-Assay: CV<12%	
Restrictions:	For Research Use only	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5657089 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Н	land	lina
11	ana	mig

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months