

# Datasheet for ABIN5657091

# **MYH3 ELISA Kit**



#### Overview

Quantity:	96 tests
Target:	MYH3
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

#### **Product Details**

Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3). No significant cross-reactivity or interference between Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3) and analogues was observed.
Sensitivity:	0.126 ng/mL

# **Target Details**

Target:	MYH3
Alternative Name:	Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3 Products)

# **Target Details** Gene Name: Myosin Heavy Chain 3, Skeletal Muscle, Embryonic Background: Gene Aliases: HEMHC, MYHC-EMB, MYHSE1, SMHCE, Myosin, Skeletal, Heavy chain, Embryonic 1, Myosin heavy chain, fast skeletal muscle, embryonic Gene ID: 4621 UniProt: P11055 **Application Details** Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end. Assay Time: 3 h Pre-coated Plate: Protocol: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3), biotinconjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzymesubstrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3) in the samples is then determined by comparing the O.D. of the samples to the standard curve.

replicates in each plate. CV(%) = SD/meanX100

respectively

Intra-Assay: CV<10%

Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level

Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level

Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3) were tested on 3 different plates, 8

Myosin Heavy Chain 3, Skeletal Muscle, Embryonic (MYH3) were tested 20 times on one plate,

Assay Precision:

# **Application Details**

	Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months