

Datasheet for ABIN5657101

MYH8 ELISA Kit



Go to Product page

Overviev	

Quantity:	96 tests
Target:	MYH8
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	62.5 pg/mL - 4000 pg/mL
Minimum Detection Limit:	62.5 pg/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8). No significant cross-reactivity or interference between Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8) and analogues was observed.
Sensitivity:	24.8 pg/mL

Target Details

Target:	MYH8
Alternative Name:	Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8 Products)

Target Details Gene Name: Myosin Heavy Chain 8, Skeletal Muscle, Perinatal Background: Gene Aliases: MyHC-peri, MyHC-pn, MyHC-perinatal, Myosin heavy chain, skeletal muscle, perinatal UniProt: P04462 **Application Details** Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end. Assay Time: 3 h Plate: Pre-coated Protocol: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8) in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Assay Precision:

replicates in each plate. CV(%) = SD/meanX100

respectively

Intra-Assay: CV<10%

Inter-Assay: CV<12%

Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8) were tested 20 times on one plate,

Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Myosin Heavy Chain 8, Skeletal Muscle, Perinatal (MYH8) were tested on 3 different plates, 8

Application Details

Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and
	specimen samples should be assayed in duplicate. Once the procedure has been started, all
	steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at
	4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant
	pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months