

## Datasheet for ABIN5657108 **MYLK ELISA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	MYLK
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

### Product Details

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Myosin Light Chain Kinase (MYLK). No significant cross-reactivity or interference between Myosin Light Chain Kinase (MYLK) and analogues was observed.
Sensitivity:	0.059 ng/mL

### Target Details

Target:	MYLK
Alternative Name:	Myosin Light Chain Kinase ( <a href="#">MYLK Products</a> )

## Target Details

---

Background: Gene Name: Myosin Light Chain Kinase  
Gene Aliases: KRP, MLCK108, MLCK210, MSTP083, MLCK, s mMLCK, MYLK1, MLCK1, Telokin, Myosin Light Chain Kinase, Smooth Muscle, deglutamylated form, Kinase-related protein

## Application Details

---

Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time: 3 h

Plate: Pre-coated

Protocol: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Myosin Light Chain Kinase (MYLK). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Myosin Light Chain Kinase (MYLK). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Myosin Light Chain Kinase (MYLK), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of  $450\text{nm} \pm 10\text{nm}$ . The concentration of Myosin Light Chain Kinase (MYLK) in the samples is then determined by comparing the O.D. of the samples to the standard curve.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Myosin Light Chain Kinase (MYLK) were tested 20 times on one plate, respectively  
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Myosin Light Chain Kinase (MYLK) were tested on 3 different plates, 8 replicates in each plate.  
 $CV(\%) = SD/\text{mean} \times 100$   
Intra-Assay:  $CV < 10\%$   
Inter-Assay:  $CV < 12\%$

Restrictions: For Research Use only

## Handling

---

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months