

Datasheet for ABIN5657211

MME ELISA Kit



Overview

Quantity:	96 tests
Target:	MME
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	31.2 pg/mL - 2000 pg/mL
Minimum Detection Limit:	31.2 pg/mL
Application:	ELISA
Product Details	
Sample Type:	Cerebrospinal Fluid, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Neprilysin (NEP). No significant cross-reactivity or interference between Neprilysin (NEP) and analogues was observed.
Sensitivity:	11.9 pg/mL
Target Details	
Target:	MME
Alternative Name:	Neprilysin (MME Products)

Target Details

•	
Target Type:	Chemical
Background:	Gene Name: Neprilysin
	Gene Aliases: CD10, MME, CALLA, SFE, Atriopeptidase, Enkephalinase, Skin fibroblast elastase,
	Common Acute Lymphoblastic Leukemia Antigen, Membrane Metallo-Endopeptidase
Gene ID:	4311
UniProt:	P08473
Pathways:	RTK Signaling, Peptide Hormone Metabolism, Regulation of Systemic Arterial Blood Pressure
	by Hormones, Smooth Muscle Cell Migration
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Neprilysin (NEP). Standard
	or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated
	antibody specific to Neprilysin (NEP). Next, Avidin conjugated to Horseradish Peroxidase (HRP)
	is added to each microplate well and incubated. After TMB substrate solution is added, only
	those wells that contain Neprilysin (NEP), biotin-conjugated antibody and enzyme-conjugated
	Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the
	addition of sulphuric acid solution and the color change is measured spectrophotometrically at
	a wavelength of 450nm ± 10nm. The concentration of Neprilysin (NEP) in the samples is then
	determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Neprilysin (NEP) were tested 20 times on one plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Neprilysin (NEP) were tested on 3 different plates, 8 replicates in each plate. CV(%) =
	SD/meanX100

Application Details

	Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months