

## Datasheet for ABIN5657310 **NPS ELISA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	NPS
Reactivity:	Human
Method Type:	Competition ELISA
Detection Range:	12.35 pg/mL - 1000 pg/mL
Minimum Detection Limit:	12.35 pg/mL
Application:	ELISA

### Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Neuropeptide S (NPS). No significant cross-reactivity or interference between Neuropeptide S (NPS) and analogues was observed.
Sensitivity:	4.48 pg/mL

### Target Details

Target:	NPS
Alternative Name:	Neuropeptide S ( <a href="#">NPS Products</a> )

## Target Details

---

Background: Gene Name: Neuropeptide S  
Gene Aliases: NP-S

Gene ID: 594857

UniProt: [P0C0P6](#)

## Application Details

---

Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time: 2 h

Plate: Pre-coated

Protocol: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Neuropeptide S (NPS) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Neuropeptide S (NPS) and unlabeled Neuropeptide S (NPS) (Standards or samples) with the pre-coated antibody specific to Neuropeptide S (NPS). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Neuropeptide S (NPS) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Neuropeptide S (NPS) in the sample.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Neuropeptide S (NPS) were tested 20 times on one plate, respectively  
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Neuropeptide S (NPS) were tested on 3 different plates, 8 replicates in each plate.  $CV(\%) = \frac{SD}{mean} \times 100$   
Intra-Assay:  $CV < 10\%$   
Inter-Assay:  $CV < 12\%$

Restrictions: For Research Use only

## Handling

---

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months