

## Datasheet for ABIN5657428 NONO ELISA Kit



Overview

Quantity:	96 tests
Target:	NONO
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

## Product Details

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Non POU Domain Containing Octamer Binding Protein (NONO). No significant cross-reactivity or interference between Non POU Domain Containing Octamer Binding Protein (NONO) and analogues was observed.
Sensitivity:	0.056 ng/mL

## Target Details

Target:	NONO	
Alternative Name:	Non POU Domain Containing Octamer Binding Protein (NONO Products)	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5657428 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target D	Details
----------	---------

Background:

Gene Name: Non POU Domain Containing Octamer Binding Protein Gene Aliases: P54, NMT55, NRB54, P54NRB, Nuclear RNA-Binding Protein 54-kD, 54 kDa nuclear RNA- and DNA-binding protein, DNA-binding p52/p100 complex, 52 kDa subunit

## **Application Details**

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than	
	5 % within the expiration date under appropriate storage condition. To minimize extra influence	
	on the performance, operation procedures and lab conditions, especially room temperature, air	
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that	
	the whole assay is performed by the same operator from the beginning to the end.	
Assay Time:	3 h	
Plate:	Pre-coated	
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate	
	provided in this kit has been pre-coated with an antibody specific to Non POU Domain	
	Containing Octamer Binding Protein (NONO). Standards or samples are then added to the	
	appropriate microtiter plate wells with a biotin-conjugated antibody specific to Non POU	
	Domain Containing Octamer Binding Protein (NONO). Next, Avidin conjugated to Horseradish	
	Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution	
	is added, only those wells that contain Non POU Domain Containing Octamer Binding Protein	
	(NONO), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in	
	color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution	
	and the color change is measured spectrophotometrically at a wavelength of 450nm $\pm$ 10nm.	
	The concentration of Non POU Domain Containing Octamer Binding Protein (NONO) in the	
	samples is then determined by comparing the O.D. of the samples to the standard curve.	
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Non	
	POU Domain Containing Octamer Binding Protein (NONO) were tested 20 times on one plate,	
	respectively	
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level	
	Non POU Domain Containing Octamer Binding Protein (NONO) were tested on 3 different	
	plates, 8 replicates in each plate. CV(%) = SD/meanX100	
	Intra-Assay: CV<10%	
	Inter-Assay: CV<12%	
Restrictions:	For Research Use only	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5657428 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Н	land	lina
11	ana	mig

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months