

## Datasheet for ABIN5657678 **PYY3 ELISA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	PYY3
Reactivity:	Human
Method Type:	Competition ELISA
Detection Range:	12.35 pg/mL - 1000 pg/mL
Minimum Detection Limit:	12.35 pg/mL
Application:	ELISA

### Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Peptide YY 3 (PYY3). No significant cross-reactivity or interference between Peptide YY 3 (PYY3) and analogues was observed.
Sensitivity:	4.82 pg/mL

### Target Details

Target:	PYY3
Alternative Name:	Peptide YY 3 ( <a href="#">PYY3 Products</a> )

## Target Details

---

Background: Gene Name: Peptide YY 3

## Application Details

---

Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time: 2 h

Plate: Pre-coated

Protocol: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Peptide YY 3 (PYY3) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Peptide YY 3 (PYY3) and unlabeled Peptide YY 3 (PYY3) (Standards or samples) with the pre-coated antibody specific to Peptide YY 3 (PYY3). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Peptide YY 3 (PYY3) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Peptide YY 3 (PYY3) in the sample.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Peptide YY 3 (PYY3) were tested 20 times on one plate, respectively  
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Peptide YY 3 (PYY3) were tested on 3 different plates, 8 replicates in each plate.  $CV(\%) = \frac{SD}{mean} \times 100$   
Intra-Assay:  $CV < 10\%$   
Inter-Assay:  $CV < 12\%$

Restrictions: For Research Use only

## Handling

---

Handling Advice: The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Storage: 4 °C, -20 °C

## Handling

---

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

---

Expiry Date: 4-8 months