

Datasheet for ABIN5657735 PPARD ELISA Kit



[Go to Product page](#)

Overview

Quantity: 96 tests

Target: PPARD

Reactivity: Human

Method Type: Sandwich ELISA

Detection Range: 0.312 ng/mL - 20 ng/mL

Minimum Detection Limit: 0.312 ng/mL

Application: ELISA

Product Details

Sample Type: Cell Culture Supernatant, Cell Lysate, Tissue Homogenate

Analytical Method: Quantitative

Detection Method: Colorimetric

Specificity: This assay has high sensitivity and excellent specificity for detection of Peroxisome Proliferator Activated Receptor Delta (PPARd). No significant cross-reactivity or interference between Peroxisome Proliferator Activated Receptor Delta (PPARd) and analogues was observed.

Sensitivity: 0.115 ng/mL

Target Details

Target: PPARD

Alternative Name: Peroxisome Proliferator Activated Receptor Delta ([PPARD Products](#))

Target Details

Background:	Gene Name: Peroxisome Proliferator Activated Receptor Delta Gene Aliases: FAAR, NR1C2, NUC1, NUCI, PPARB, PPARd/B, Nuclear Receptor Subfamily 1, Group C, Member 2, Nuclear hormone receptor 1, Peroxisome proliferator-activated receptor beta
Gene ID:	5467
UniProt:	Q03181
Pathways:	Nuclear Receptor Transcription Pathway , Positive Regulation of Peptide Hormone Secretion , Steroid Hormone Mediated Signaling Pathway , Monocarboxylic Acid Catabolic Process , Smooth Muscle Cell Migration , Positive Regulation of fat Cell Differentiation

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Peroxisome Proliferator Activated Receptor Delta (PPARd). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Peroxisome Proliferator Activated Receptor Delta (PPARd). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Peroxisome Proliferator Activated Receptor Delta (PPARd), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of Peroxisome Proliferator Activated Receptor Delta (PPARd) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Peroxisome Proliferator Activated Receptor Delta (PPARd) were tested 20 times on one plate, respectively

Application Details

Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Peroxisome Proliferator Activated Receptor Delta (PPAR δ) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100
Intra-Assay: CV<10%
Inter-Assay: CV<12%

Restrictions: For Research Use only

Handling

Handling Advice: The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Storage: 4 °C,-20 °C

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months
