

Datasheet for ABIN5657829 **PIM1 CLIA Kit**



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	PIM1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.781 ng/mL - 50 ng/mL
Minimum Detection Limit:	0.781 ng/mL
Application:	ELISA

Product Details

Sample Type:	Cell Culture Supernatant, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Specificity:	This assay has high sensitivity and excellent specificity for detection of Pim-1 Oncogene (PIM1). No significant cross-reactivity or interference between Pim-1 Oncogene (PIM1) and analogues was observed.
Sensitivity:	0.03 ng/mL

Target Details

Target:	PIM1
Alternative Name:	Pim-1 Oncogene (PIM1 Products)

Target Details

Background:	Gene Name: Pim-1 Oncogene Gene Aliases: Proto-Oncogene Serine/Threonine-Protein Kinase Pim-1
Gene ID:	5292
UniProt:	P11309
Pathways:	Glycosaminoglycan Metabolic Process

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 - 3 h
Plate:	Pre-coated
Protocol:	The microplate provided in this kit has been pre-coated with an antibody specific to Pim-1 Oncogene (PIM1). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Pim-1 Oncogene (PIM1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Pim-1 Oncogene (PIM1) level in the sample or standard.,
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Pim-1 Oncogene (PIM1) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Pim-1 Oncogene (PIM1) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only

Handling

Handling Advice:	Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be
------------------	--

Handling

assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Storage: 4 °C, -20 °C

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months