

### Datasheet for ABIN5658082

### **PDCD5 ELISA Kit**



_				
( )	VE	r\/		Λ/
( )	V C	I V	1	v v

Quantity:	96 tests
Target:	PDCD5
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

#### **Product Details**

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Programmed Cell Death Protein 5 (PDCD5). No significant cross-reactivity or interference between Programmed Cell Death Protein 5 (PDCD5) and analogues was observed.
Sensitivity:	0.057 ng/mL

### Target Details

Target:	PDCD5
Alternative Name: Programmed Cell Death Protein 5 (PDCD5 Products)	

## Target Details

Restrictions:

larget Details	
Background:	Gene Name: Programmed Cell Death Protein 5
	Gene Aliases: TFAR19, TFAR19 Novel Apoptosis-Related, TF1 Cell Apoptosis-Related Gene 19
Gene ID:	9141
UniProt:	014737
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Programmed Cell Death
	Protein 5 (PDCD5). Standards or samples are then added to the appropriate microtiter plate
	wells with a biotin-conjugated antibody specific to Programmed Cell Death Protein 5 (PDCD5).
	Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and
	incubated. After TMB substrate solution is added, only those wells that contain Programmed
	Cell Death Protein 5 (PDCD5), biotin-conjugated antibody and enzyme-conjugated Avidin will
	exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of
	sulphuric acid solution and the color change is measured spectrophotometrically at a
	wavelength of 450nm ± 10nm. The concentration of Programmed Cell Death Protein 5 (PDCD5
	in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Programmed Cell Death Protein 5 (PDCD5) were tested 20 times on one plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Programmed Cell Death Protein 5 (PDCD5) were tested on 3 different plates, 8 replicates in
	each plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%
	Inter-Assay: CV<12%
<b>5</b>	

For Research Use only

# Handling

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months