

Datasheet for ABIN5658128 **PCSK1 ELISA Kit**



[Go to Product page](#)

Overview

Quantity: 96 tests

Target: PCSK1

Reactivity: Rat

Method Type: Sandwich ELISA

Detection Range: 3.12 ng/mL - 200 ng/mL

Minimum Detection Limit: 3.12 ng/mL

Application: ELISA

Product Details

Sample Type: Plasma, Serum

Analytical Method: Quantitative

Detection Method: Colorimetric

Specificity: This assay has high sensitivity and excellent specificity for detection of Proprotein Convertase Subtilisin/Kexin Type 1 (PCSK1). No significant cross-reactivity or interference between Proprotein Convertase Subtilisin/Kexin Type 1 (PCSK1) and analogues was observed.

Sensitivity: 1.44 ng/mL

Target Details

Target: PCSK1

Alternative Name: Proprotein Convertase Subtilisin/Kexin Type 1 ([PCSK1 Products](#))

Target Details

Background: Gene Name: Proprotein Convertase Subtilisin/Kexin Type 1
Gene Aliases: PC1, PC3, SPC3, NEC1, Prohormone Convertase 3, Prohormone Convertase 1, Neuroendocrine Convertase 1, Proprotein Convertase 1

Pathways: [Peptide Hormone Metabolism](#)

Application Details

Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time: 3 h

Plate: Pre-coated

Protocol: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Proprotein Convertase Subtilisin/Kexin Type 1 (PCSK1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Proprotein Convertase Subtilisin/Kexin Type 1 (PCSK1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Proprotein Convertase Subtilisin/Kexin Type 1 (PCSK1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of Proprotein Convertase Subtilisin/Kexin Type 1 (PCSK1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Proprotein Convertase Subtilisin/Kexin Type 1 (PCSK1) were tested 20 times on one plate, respectively
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Proprotein Convertase Subtilisin/Kexin Type 1 (PCSK1) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100
Intra-Assay: CV<10%
Inter-Assay: CV<12%

Application Details

Restrictions: For Research Use only

Handling

Handling Advice: The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Storage: 4 °C,-20 °C

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months