

## Datasheet for ABIN5658314 PDHB ELISA Kit



Overview

Quantity:	96 tests
Target:	PDHB
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	1.56 ng/mL - 100 ng/mL
Minimum Detection Limit:	1.56 ng/mL
Application:	ELISA

## Product Details

Sample Type:	Cell Culture Supernatant, Cell Lysate, Tissue Homogenate	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of Pyruvate Dehydrogenase Beta (PDHb). No significant cross-reactivity or interference between Pyruvate Dehydrogenase Beta (PDHb) and analogues was observed.	
Sensitivity:	0.65 ng/mL	
Target Details		
Target:	PDHB	

Alternative Name:	Pyruvate Dehydrogenase Beta (PDHB Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5658314 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details			
Background:	Gene Name: Pyruvate Dehydrogenase Beta		
	Gene Aliases: PDH-B, PHE1B, PHE1-B, Pyruvate Dehydrogenase(lipoamide)Beta		
Gene ID:	5162		
UniProt:	P11177		
Pathways:	Warburg Effect		
Application Details			
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.		
Assay Time:	3 h		
Plate:	Pre-coated		
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Pyruvate Dehydrogenase Beta (PDHb). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Pyruvate Dehydrogenase Beta (PDHb). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Pyruvate Dehydrogenase Beta (PDHb), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Pyruvate Dehydrogenase Beta (PDHb) in the samples is then determined by comparing the 0.D. of the samples to the standard curve.		
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Pyruvate Dehydrogenase Beta (PDHb) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Pyruvate Dehydrogenase Beta (PDHb) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%		
Restrictions:	For Research Use only		

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5658314 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Н	land	lina
11	ana	mig

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months