

Datasheet for ABIN5658336 **RAB1A CLIA Kit**



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	RAB1A
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	156.25 pg/mL - 10000 pg/mL
Minimum Detection Limit:	156.25 pg/mL
Application:	ELISA

Product Details

Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Specificity:	This assay has high sensitivity and excellent specificity for detection of RAB1A, Member RAS Oncogene Family (RAB1A). No significant cross-reactivity or interference between RAB1A, Member RAS Oncogene Family (RAB1A) and analogues was observed.
Sensitivity:	5.27 pg/mL

Target Details

Target:	RAB1A
Alternative Name:	RAB1A, Member RAS Oncogene Family (RAB1A Products)

Target Details

Background:	Gene Name: RAB1A, Member RAS Oncogene Family Gene Aliases: RAB1, YPT1, Ras-related protein Rab-1A, YPT1-related protein
Gene ID:	5861
UniProt:	P62820
Pathways:	SARS-CoV-2 Protein Interactome

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 - 3 h
Plate:	Pre-coated
Protocol:	The microplate provided in this kit has been pre-coated with an antibody specific to RAB1A, Member RAS Oncogene Family (RAB1A). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to RAB1A, Member RAS Oncogene Family (RAB1A). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the RAB1A, Member RAS Oncogene Family (RAB1A) level in the sample or standard.,
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level RAB1A, Member RAS Oncogene Family (RAB1A) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level RAB1A, Member RAS Oncogene Family (RAB1A) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Restrictions:	For Research Use only

Handling

Handling Advice:	Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months