

Datasheet for ABIN5658719

SPTLC1 ELISA Kit



Go to Product page

_					
	W	0	rv	10	W

Quantity:	96 tests
Target:	SPTLC1
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

Product Details

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1). No significant cross-reactivity or interference between Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1) and analogues was observed.
Sensitivity:	0.054 ng/mL

Target Details

Target:	SPTLC1
Alternative Name:	Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1 Products)

Target Details Background: Gene ID: UniProt: Application Details

Gene Name: Serine Palmitoyltransferase, Long Chain Base Subunit 1

Gene Aliases: HSAN, HSAN1, HSN1, LBC1, LCB1, SPT1, SPTI, Hereditary Sensory

Neuropathy, Type 1, Long chain base biosynthesis protein 1, Serine-palmitoyl-CoA transferase 1

268656

035704

Comment:

The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time:

3 h

Plate:

Pre-coated

Protocol:

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.

Assay Precision:

Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1) were tested 20 times on one plate, respectively

Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Serine Palmitoyltransferase, Long Chain Base Subunit 1 (SPTLC1) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100

Application Details

	Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months