

Datasheet for ABIN5658773 SH3BGRL ELISA Kit



o to i roudot page

Overview

Quantity:	96 tests
Target:	SH3BGRL
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

Product Details

Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of SH3 Domain Binding Glutamic Acid Rich Protein Like Protein (SH3BGRL). No significant cross-reactivity or interference between SH3 Domain Binding Glutamic Acid Rich Protein Like Protein (SH3BGRL) and analogues was observed.
Sensitivity:	0.061 ng/mL

Target Details

Target:	SH3BGRL
Alternative Name:	SH3 Domain Binding Glutamic Acid Rich Protein Like Protein (SH3BGRL Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5658773 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details		
Background:	Gene Name: SH3 Domain Binding Glutamic Acid Rich Protein Like Protein	
Gene ID:	6451	
UniProt:	075368	
Application Details		
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.	
Assay Time:	3 h	
Plate:	Pre-coated	
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to SH3 Domain Binding Glutamic Acid Rich Protein Like Protein (SH3BGRL). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to SH3 Domain Binding Glutamic Acid Rich Protein Like Protein (SH3BGRL). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain SH3 Domain Binding Glutamic Acid Rich Protein (SH3BGRL), biotin-conjugated antibody and enzyme-conjugated Avidir will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of SH3 Domain Binding Glutamic Acid Rich Protein Like Protein (SH3BGRL) in the samples is then determined by comparing the O.D. of the samples to the standard curve.	
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level SH Domain Binding Glutamic Acid Rich Protein Like Protein (SH3BGRL) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level SH3 Domain Binding Glutamic Acid Rich Protein Like Protein (SH3BGRL) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5658773 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Application Details	
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months