

Datasheet for ABIN5659660 TNFSF9 ELISA Kit



Overview

Quantity:	96 tests
Target:	TNFSF9
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	7.81 pg/mL - 500 pg/mL
Minimum Detection Limit:	7.81 pg/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Tumor Necrosis Factor Ligand Superfamily, Member 9 (TNFSF9). No significant cross-reactivity or interference between Tumor Necrosis Factor Ligand Superfamily, Member 9 (TNFSF9) and analogues was observed.
Sensitivity:	2.5 pg/mL
Target Details	

Target:	TNFSF9
Alternative Name:	Tumor Necrosis Factor Ligand Superfamily, Member 9 (TNFSF9 Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5659660 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details	
Background:	Gene Name: Tumor Necrosis Factor Ligand Superfamily, Member 9
	Gene Aliases: 4-1BB-L, Receptor 4-1BB Ligand, Homolog Of Mouse 4-1BB-L
Gene ID:	8744
UniProt:	P41273
Pathways:	Activated T Cell Proliferation, Cancer Immune Checkpoints
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Tumor Necrosis Factor
	Ligand Superfamily, Member 9 (TNFSF9). Standards or samples are then added to the
	appropriate microtiter plate wells with a biotin-conjugated antibody specific to Tumor Necrosis
	Factor Ligand Superfamily, Member 9 (TNFSF9). Next, Avidin conjugated to Horseradish
	Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution
	is added, only those wells that contain Tumor Necrosis Factor Ligand Superfamily, Member 9
	(TNFSF9), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in
	color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution
	and the color change is measured spectrophotometrically at a wavelength of 450nm \pm 10nm.
	The concentration of Tumor Necrosis Factor Ligand Superfamily, Member 9 (TNFSF9) in the
	samples is then determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Tumor Necrosis Factor Ligand Superfamily, Member 9 (TNFSF9) were tested 20 times on one
	plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Tumor Necrosis Factor Ligand Superfamily, Member 9 (TNFSF9) were tested on 3 different
	plates, 8 replicates in each plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5659660 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Application Details		
	Inter-Assay: CV<12%	
Restrictions:	For Research Use only	
Handling		
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.	
Storage:	4 °C,-20 °C	
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.	
Expiry Date:	4-8 months	