

## Datasheet for ABIN5659994

## **WNT1 ELISA Kit**



#### Overview

Quantity:	96 tests
Target:	WNT1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

### **Product Details**

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Wingless Type MMTV Integration Site Family, Member 1 (WNT1). No significant cross-reactivity or interference between Wingless Type MMTV Integration Site Family, Member 1 (WNT1) and analogues was observed.
Sensitivity:	0.057 ng/mL

## Target Details

Target:	WNT1
Alternative Name:	Wingless Type MMTV Integration Site Family, Member 1 (WNT1 Products)

Target Details	
Background:	Gene Name: Wingless Type MMTV Integration Site Family, Member 1
	Gene Aliases: INT1, Proto-oncogene Int-1 homolog
Pathways:	WNT Signaling, Dopaminergic Neurogenesis
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5% within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Wingless Type MMTV
	Integration Site Family, Member 1 (WNT1). Standards or samples are then added to the
	appropriate microtiter plate wells with a biotin-conjugated antibody specific to Wingless Type
	MMTV Integration Site Family, Member 1 (WNT1). Next, Avidin conjugated to Horseradish
	Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution
	is added, only those wells that contain Wingless Type MMTV Integration Site Family, Member 1
	(WNT1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in
	color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution
	and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm.
	The concentration of Wingless Type MMTV Integration Site Family, Member 1 (WNT1) in the
	samples is then determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Wingless Type MMTV Integration Site Family, Member 1 (WNT1) were tested 20 times on one
	plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Wingless Type MMTV Integration Site Family, Member 1 (WNT1) were tested on 3 different
	plates, 8 replicates in each plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%
	Inter-Assay: CV<12%
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high lead Wingless Type MMTV Integration Site Family, Member 1 (WNT1) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10%

For Research Use only

Restrictions:

# Handling

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months