

[Go to Product page](#)

## Datasheet for ABIN5660037 **WISP2 CLIA Kit**

### Overview

Quantity:	96 tests
Target:	WISP2
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	312.5 pg/mL - 20000 pg/mL
Minimum Detection Limit:	312.5 pg/mL
Application:	ELISA

### Product Details

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Specificity:	This assay has high sensitivity and excellent specificity for detection of WNT1 Inducible Signaling Pathway Protein 2 (WISP2). No significant cross-reactivity or interference between WNT1 Inducible Signaling Pathway Protein 2 (WISP2) and analogues was observed.
Sensitivity:	10.3 pg/mL

### Target Details

Target:	WISP2
Alternative Name:	WNT1 Inducible Signaling Pathway Protein 2 ( <a href="#">WISP2 Products</a> )

## Target Details

---

Background:	Gene Name: WNT1 Inducible Signaling Pathway Protein 2 Gene Aliases: CCN5, CT58, CTGF-L, CCN family member 5, Connective tissue growth factor-like protein, Connective tissue growth factor-related protein 58
Gene ID:	8839
UniProt:	<a href="#">O76076</a>
Pathways:	<a href="#">WNT Signaling</a> , <a href="#">Growth Factor Binding</a>

## Application Details

---

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 - 3 h
Plate:	Pre-coated
Protocol:	The microplate provided in this kit has been pre-coated with an antibody specific to WNT1 Inducible Signaling Pathway Protein 2 (WISP2). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to WNT1 Inducible Signaling Pathway Protein 2 (WISP2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the WNT1 Inducible Signaling Pathway Protein 2 (WISP2) level in the sample or standard.,
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level WNT1 Inducible Signaling Pathway Protein 2 (WISP2) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level WNT1 Inducible Signaling Pathway Protein 2 (WISP2) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only

## Handling

---

Handling Advice:	Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months