

Datasheet for ABIN5660085 **ZG16B CLIA Kit**



[Go to Product page](#)

Overview

| | |
|--------------------------|---------------------------|
| Quantity: | 96 tests |
| Target: | ZG16B |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 312.5 pg/mL - 20000 pg/mL |
| Minimum Detection Limit: | 312.5 pg/mL |
| Application: | ELISA |

Product Details

| | |
|--------------------|---|
| Sample Type: | Plasma, Serum |
| Analytical Method: | Quantitative |
| Detection Method: | Chemiluminescent |
| Specificity: | This assay has high sensitivity and excellent specificity for detection of Zymogen Granule Protein 16 Homolog B (ZG16B). No significant cross-reactivity or interference between Zymogen Granule Protein 16 Homolog B (ZG16B) and analogues was observed. |
| Sensitivity: | 11.16 pg/mL |

Target Details

| | |
|-------------------|---|
| Target: | ZG16B |
| Alternative Name: | Zymogen Granule Protein 16 Homolog B (ZG16B Products) |

Target Details

| | |
|-------------|---|
| Background: | Gene Name: Zymogen Granule Protein 16 Homolog B Gene Aliases: JCLN2, Jacalin-Like Lectin Domain Containing 2 |
| Gene ID: | 124220 |
| UniProt: | Q96DA0 |

Application Details

| | |
|------------------|--|
| Comment: | The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end. |
| Assay Time: | 2 - 3 h |
| Plate: | Pre-coated |
| Protocol: | The microplate provided in this kit has been pre-coated with an antibody specific to Zymogen Granule Protein 16 Homolog B (ZG16B). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Zymogen Granule Protein 16 Homolog B (ZG16B). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Zymogen Granule Protein 16 Homolog B (ZG16B) level in the sample or standard., |
| Assay Precision: | Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Zymogen Granule Protein 16 Homolog B (ZG16B) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Zymogen Granule Protein 16 Homolog B (ZG16B) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$ |
| Restrictions: | For Research Use only |

Handling

| | |
|------------------|--|
| Handling Advice: | Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption. |
| Storage: | 4 °C,-20 °C |
| Storage Comment: | -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles. |
| Expiry Date: | 4-8 months |