

Datasheet for ABIN5668748

**SORD CLIA Kit**[Go to Product page](#)**1** Image

## Overview

Quantity:	96 tests
Target:	SORD
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.625 ng/mL - 40 ng/mL
Minimum Detection Limit:	0.625 ng/mL
Application:	ELISA

## Product Details

Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Sensitivity:	0.375 ng/mL
Components:	<ul style="list-style-type: none"><li>- Micro CLIA Plate</li><li>- Reference Standard</li><li>- Concentrated Biotinylated Detection Ab (100×)</li><li>- Concentrated HRP Conjugate (100×)</li><li>- Reference Standard &amp; Sample Diluent</li><li>- Biotinylated Detection Ab Diluent</li><li>- HRP Conjugate Diluent</li><li>- Concentrated Wash Buffer (25×)</li><li>- Substrate Reagent A</li><li>- Substrate Reagent B</li></ul>

## Product Details

- Plate Sealer
- Product Description
- Certificate of Analysis

## Target Details

Target:	SORD
Alternative Name:	Sorbitol Dehydrogenase ( <a href="#">SORD Products</a> )


## Application Details

Plate:	Pre-coated
Restrictions:	For Research Use only

## Handling

Storage:	4 °C, -20 °C
Storage Comment:	The unopened kit can be stored at 4 °C for 1 week. If the kit is not used within 1 week, store the items separately according to the special conditions since the kit is received.

## Images



### ELISA

**Image 1.** Typical standard curve