

Datasheet for ABIN5674708
NUCB2 ELISA Kit



[Go to Product page](#)

Overview

| | |
|--------------------------|--------------------------|
| Quantity: | 96 tests |
| Target: | NUCB2 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 31.25 pg/mL - 2000 pg/mL |
| Minimum Detection Limit: | 31.25 pg/mL |
| Application: | ELISA |

Product Details

| | |
|--------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Sample Type: | Plasma, Serum |
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Sensitivity: | 13.8 pg/mL |
| Components: | Pre-coated, ready to use 96-well strip plate Standard (freeze dried) Standard Diluent Detection Reagent A Detection Reagent B Assay Diluent A Assay Diluent B TMB Substhumane |

Product Details

Stop Solution
Wash Buffer(30 x concenthumane)
Plate sealer for 96 wells
Instruction manual

Material not included:

1. Microplate reader with $450 \pm 10\text{nm}$ filter.
2. Precision single or multi-channel pipettes and disposable tips.
3. Eppendorf Tubes for diluting samples.
4. Deionized or distilled water.
5. Absorbent paper for blotting the microtiter plate.
6. Container for Wash Solution.

Target Details

Target: NUCB2

Abstract: [NUCB2 Products](#)

Background: Alternative name: NUCB2, NEFA, Nucleobindin 2, DNA-binding protein NEFA, Gastric cancer antigen Zg4, Prepronesfatin

Gene ID: 4925

UniProt: [P80303](#)

Application Details

Application Notes: Optimal working dilution should be determined by the investigator.

Assay Time: 1 - 4.5 h

Plate: Pre-coated

Protocol:

1. Prepare all reagents, samples and standards
2. Add 100 μL standard or sample to each well. Incubate 2 hours at 37°C
3. Aspirate and add 100 μL prepared Detection Reagent A. Incubate 1 hour at 37°C
4. Aspirate and wash 3 times
5. Add 100 μL prepared Detection Reagent B. Incubate 1 hour at 37°C
6. Aspirate and wash 5 times
7. Add 90 μL Substrate Solution. Incubate 15-25 minutes at 37°C
8. Add 50 μL Stop Solution. Read at 450nm immediately.

Restrictions: For Research Use only

Handling

Storage: 4 °C

Expiry Date: 12 months