

Datasheet for ABIN5674723
SOCS2 ELISA Kit



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	SOCS2
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

Product Details

Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Sensitivity:	0059 ng/mL
Components:	Pre-coated, ready to use 96-well strip plate Standard (freeze dried) Standard Diluent Detection Reagent A Detection Reagent B Assay Diluent A Assay Diluent B TMB Substhumane

Product Details

Stop Solution
Wash Buffer(30 x concenthumane)
Plate sealer for 96 wells
Instruction manual

Material not included:

1. Microplate reader with 450 ± 10 nm filter.
2. Precision single or multi-channel pipettes and disposable tips.
3. Eppendorf Tubes for diluting samples.
4. Deionized or distilled water.
5. Absorbent paper for blotting the microtiter plate.
6. Container for Wash Solution.

Target Details

Target: SOCS2

Alternative Name: Suppressors Of Cytokine Signaling 2 (SOCS2) ([SOCS2 Products](#))

Background: Alternative name: CIS2, Cish2, SOCS-2, SSI-2, SSI2, STATI2, Cytokine-inducible SH2 protein 2, STAT-induced STAT inhibitor 2

Gene ID: 8835

UniProt: [O14508](#)

Pathways: [JAK-STAT Signaling](#), [Response to Growth Hormone Stimulus](#)

Application Details

Application Notes: Optimal working dilution should be determined by the investigator.

Assay Time: 1 - 4.5 h

Plate: Pre-coated

Protocol:

1. Prepare all reagents, samples and standards
2. Add 100 μ L standard or sample to each well. Incubate 2 hours at 37°C
3. Aspirate and add 100 μ L prepared Detection Reagent A. Incubate 1 hour at 37°C
4. Aspirate and wash 3 times
5. Add 100 μ L prepared Detection Reagent B. Incubate 1 hour at 37°C
6. Aspirate and wash 5 times
7. Add 90 μ L Substrate Solution. Incubate 15-25 minutes at 37°C
8. Add 50 μ L Stop Solution. Read at 450nm immediately.

Application Details

Restrictions: For Research Use only

Handling

Storage: 4 °C

Expiry Date: 12 months