

Datasheet for ABIN5674747  
**APEX1 ELISA Kit**



[Go to Product page](#)

## Overview

|                          |                        |
|--------------------------|------------------------|
| Quantity:                | 96 tests               |
| Target:                  | APEX1                  |
| Reactivity:              | Human                  |
| Method Type:             | Sandwich ELISA         |
| Detection Range:         | 0.312 ng/mL - 20 ng/mL |
| Minimum Detection Limit: | 0.312 ng/mL            |
| Application:             | ELISA                  |

## Product Details

|                    |  |
|--------------------|--|
| Sample Type:       | Cell Lysate, Tissue Homogenate   |
| Analytical Method: | Quantitative   |
| Detection Method:  | Colorimetric   |
| Sensitivity:       | 0132 ng/mL   |
| Components:        | Pre-coated, ready to use 96-well strip plate<br>Standard (freeze dried)<br>Standard Diluent<br>Detection Reagent A<br>Detection Reagent B<br>Assay Diluent A<br>Assay Diluent B<br>TMB Substhumane |

## Product Details

---

Stop Solution  
Wash Buffer(30 x concenthumane)  
Plate sealer for 96 wells  
Instruction manual

---

Material not included:

1. Microplate reader with  $450 \pm 10\text{nm}$  filter.
2. Precision single or multi-channel pipettes and disposable tips.
3. Eppendorf Tubes for diluting samples.
4. Deionized or distilled water.
5. Absorbent paper for blotting the microtiter plate.
6. Container for Wash Solution.

## Target Details

---

Target: APEX1

---

Alternative Name: APEX Nuclease 1 (APEX1) ([APEX1 Products](#))

---

Background: Alternative name: APE, APE1, APEN, APEX, APX, HAP1, REF1, Multifunctional DNA Repair Enzyme, APEX Nuclease 1, Redox factor-1, DNA-(apurinic or apyrimidinic site) lyase

---

Gene ID: 328

---

UniProt: [P27695](#)

---

Pathways: [DNA Damage Repair](#), [Chromatin Binding](#), [Cell RedoxHomeostasis](#), [Smooth Muscle Cell Migration](#), [Positive Regulation of Response to DNA Damage Stimulus](#)

## Application Details

---

Application Notes: Optimal working dilution should be determined by the investigator.

---

Assay Time: 1 - 4.5 h

---

Plate: Pre-coated

---

Protocol:

1. Prepare all reagents, samples and standards
2. Add 100 $\mu\text{L}$  standard or sample to each well. Incubate 2 hours at 37°C
3. Aspirate and add 100 $\mu\text{L}$  prepared Detection Reagent A. Incubate 1 hour at 37°C
4. Aspirate and wash 3 times
5. Add 100 $\mu\text{L}$  prepared Detection Reagent B. Incubate 1 hour at 37°C
6. Aspirate and wash 5 times
7. Add 90 $\mu\text{L}$  Substrate Solution. Incubate 15-25 minutes at 37°C

---

## Application Details

---

8. Add 50 $\mu$ L Stop Solution. Read at 450nm immediately.

---

Restrictions: For Research Use only

## Handling

---

Storage: 4 °C

---

Expiry Date: 12 months