

Datasheet for ABIN5690771
CD40 Ligand ELISA Kit



[Go to Product page](#)

Overview

Quantity: 96 tests

Target: CD40 Ligand (CD40LG)

Reactivity: Dog

Method Type: Sandwich ELISA

Application: ELISA

Product Details

Purpose: Canine (dog) CD40 Ligand/TNFSF5 ELISA Kit.

Sample Type: Cell Culture Supernatant, Cell Samples, Plasma, Serum, Tissue Lysate

Analytical Method: Quantitative

Detection Method: Colorimetric

Specificity: This ELISA antibody pair recognizes Canine (dog) CD40 Ligand.

Characteristics:

- Strip plates and additional reagents allow for use in multiple experiments
- Quantitative protein detection
- Establishes normal range
- The best products for confirmation of antibody array data

Components:

- Pre-Coated 96-well Strip Microplate
- Wash Buffer
- Stop Solution
- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody

Product Details

- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µL to 1 µL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	CD40 Ligand (CD40LG)
Alternative Name:	CD40 Ligand (TNFSF5 (CD40LG Products))
Gene ID:	403468
UniProt:	O97626
Pathways:	NF-kappaB Signaling , Production of Molecular Mediator of Immune Response , Cancer Immune Checkpoints

Application Details

Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual. 2. Add 100 µL of standard or sample to each well. 3. Incubate 2.5 h at RT or O/N at 4 °C. 4. Add 100 µL of prepared biotin antibody to each well. 5. Incubate 1 h at RT. 6. Add 100 µL of prepared Streptavidin solution to each well. 7. Incubate 45 min at RT. 8. Add 100 µL of TMB One-Step Substrate Reagent to each well. 9. Incubate 30 min at RT. 10. Add 50 µL of Stop Solution to each well. 11. Read at 450 nm immediately.
Restrictions:	For Research Use only

Handling

Storage: 4 °C

Expiry Date: 6 months