

[Go to Product page](#)

Datasheet for ABIN5691007 IL17F ELISA Kit

Overview

Quantity:	96 tests
Target:	IL17F
Reactivity:	Rabbit
Method Type:	Sandwich ELISA
Application:	ELISA

Product Details

Purpose:	Rabbit IL-17F ELISA Kit.
Sample Type:	Cell Culture Supernatant, Cell Samples, Plasma, Serum, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair recognizes Rabbit IL-17F.
Characteristics:	<ul style="list-style-type: none">• Strip plates and additional reagents allow for use in multiple experiments• Quantitative protein detection• Establishes normal range• The best products for confirmation of antibody array data
Components:	<ul style="list-style-type: none">• Pre-Coated 96-well Strip Microplate• Wash Buffer• Stop Solution• Assay Diluent(s)• Lyophilized Standard• Biotinylated Detection Antibody

Product Details

- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µL to 1 µL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	IL17F
Alternative Name:	IL-17F (IL17F Products)
Gene ID:	100339570
UniProt:	G1SLF3
Pathways:	Cellular Response to Molecule of Bacterial Origin, Positive Regulation of Endopeptidase Activity

Application Details

Plate:	Pre-coated
Protocol:	<ol style="list-style-type: none">1. Prepare all reagents, samples and standards as instructed in the manual.2. Add 100 µL of standard or sample to each well.3. Incubate 2.5 h at RT or O/N at 4 °C.4. Add 100 µL of prepared biotin antibody to each well.5. Incubate 1 h at RT.6. Add 100 µL of prepared Streptavidin solution to each well.7. Incubate 45 min at RT.8. Add 100 µL of TMB One-Step Substrate Reagent to each well.9. Incubate 30 min at RT.10. Add 50 µL of Stop Solution to each well.11. Read at 450 nm immediately.
Restrictions:	For Research Use only

Handling

Storage: 4 °C

Expiry Date: 6 months