

## Datasheet for ABIN578626

# **Granulin ELISA Kit**

1 Image 2 Publications



### Overview

Quantity:	96 tests
Target:	Granulin (GRN)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	4.7-300 ng/mL
Minimum Detection Limit:	4.7 ng/mL
Application:	ELISA

Product Details	
Sample Type:	Serum, Plasma, Cell Culture Supernatant
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Cross-Reactivity (Details):	Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between the target antigen and all analogues for other species. Therefore, cross reaction may still exist.
Sensitivity:	1.17 ng/mL
Components:	<ul> <li>Assay plate (12 × 8 coated Microwells)</li> <li>Standard (freeze dried)</li> <li>Biotin-antibody (100 × concentrate)</li> <li>HRP-avidin (100 × concentrate)</li> <li>Biotin-antibody Diluent</li> </ul>

· HRP-avidin Diluent

- · Sample Diluent
- Wash Buffer (25 × concentrate)
- TMB Substrate
- · Stop Solution
- · Adhesive Strip (for 96 wells)
- · Instruction manual

Target Details	
Target:	Granulin (GRN)
Alternative Name:	Progranulin (PRGN) (GRN Products)
Target Type:	Viral Protein
Application Details	
Application Notes:	<ul> <li>The supplier is only responsible for the kit itself, but not for the samples consumed during the assay. The user should calculate the possible amount of the samples used in the whole test. Please reserve sufficient samples in advance.</li> <li>Samples to be used within 5 days may be stored at 2-8°C, otherwise samples must be stored at -20°C (≤ 1 month) or -80°C (≤ 2 months) to avoid loss of bioactivity and contamination.</li> <li>Grossly hemolyzed samples are not suitable for use in this assay.</li> <li>If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.</li> <li>Please predict the concentration before assaying. If values for these are not within the range of the standard curve, users must determine the optimal sample dilutions for their particular experiments.</li> <li>Tissue or cell extraction samples prepared by chemical lysis buffer may cause unexpected ELISA results due to the impacts of certain chemicals.</li> <li>Owing to the possibility of mismatching between antigens from another resource and antibodies used in this supplier's kits (e.g., antibody targets conformational epitope rather</li> </ul>

than linear epitope), some native or recombinant proteins from other manufacturers may not

· Influenced by factors including cell viability, cell number and cell sampling time, samples

• Fresh samples without long time storage are recommended for the test. Otherwise, protein degradation and denaturalization may occur in those samples and finally lead to wrong

Comment:

Detection wavelength: 450 nm

results.

be recognized by this supplier's products.

Information on standard material:

from cell culture supernatant may not be recognized by the kit.

Depending on the antigen to be detected, standards can be either native or recombinant protein. The recombinant proteins are being expressed in CHO cells in most cases. Please inquire for more information. The formulation of auxiliary material in the standard is considered proprietary information, however it does not contain any poisonous substance. Proclin 300 (1:3000) is used as preservative.

### Information on reagents:

In most cases the stop solution provided is 1 N H2SO4. The formulation of wash solution is proprietary information. None of the components contain (sodium) azide, thimerosal, 2-mercaptoethanol (2-ME) or any other poisonous materials. For the sandwich method kits, the sample diluent, antibody diluent, enzyme diluent and standard all contain BSA.

#### Information on antibodies:

The antibodies provided in different kits vary in regards to clonality and host. Some antibodies are affinity purified, some are Protein A

Sample Volume:	100 μL
Assay Time:	1 - 4.5 h
Plate:	Pre-coated
Assay Precision:	Intra-assay precision (precision within an assay): Three samples of known concentration were tested twenty times on one plate to assess precision.  Inter-assay precision (precision between assays): Three samples of known concentration were tested in twenty assays to assess precision.  Intra-assay: CV% less than 8%  Inter-assay: CV% less than 10%
Restrictions:	For Research Use only
Handling	
Precaution of Use:	The Stop Solution provided with this kit is an acid solution. Wear eye, hand, face and clothing

# Handling Advice:

protection when using this material.

## • The kit should not be used beyond the expiration date on the kit label.

- Do not mix or substitute reagents with those from other lots or sources.
- If samples generate values higher than the highest standard, dilute the samples with Sample Diluent and repeat the assay.

- Any variation in Sample Diluent, operator, pipetting technique, washing technique, incubation time/temperature and kit age can cause variation in binding.
- This assay is designed to eliminate interference by soluble receptors, binding proteins and other factors present in biological samples. Until all factors have been tested in the Immunoassay, the possibility of interference cannot be excluded.

Storage:	4 °C/-20 °C
Storage Comment:	For unopened kit: All the reagents should be kept according to the labels on vials.
Expiry Date:	6 months


### **Publications**

Product cited in:

Qu, Wang, Deng, Wei, Deng: "Associations between longer habitual day napping and non-alcoholic fatty liver disease in an elderly Chinese population." in: **PLoS ONE**, Vol. 9, Issue 8, pp. e105583, (2014) (PubMed).

Qu, Deng, Hu: "Plasma progranulin concentrations are increased in patients with type 2 diabetes and obesity and correlated with insulin resistance." in: **Mediators of inflammation**, Vol. 2013, pp. 360190, (2013) (PubMed).

### **Images**



### **ELISA**

Image 1. Typical standard curve