

Datasheet for ABIN6200067 **Relaxin CLIA Kit**

[Go to Product page](#)

Overview

| | |
|--------------------------|--------------------------|
| Quantity: | 96 tests |
| Target: | Relaxin |
| Reactivity: | Rabbit |
| Method Type: | Competition ELISA |
| Detection Range: | 24.69 pg/mL - 2000 pg/mL |
| Minimum Detection Limit: | 24.69 pg/mL |
| Application: | ELISA |

Product Details

| | |
|--------------------|---|
| Purpose: | Rabbit Relaxin (RLN) CLIA Kit is a Competitive CLIA Kit for use with Serum, plasma and other biological fluids. |
| Sample Type: | Plasma, Serum |
| Analytical Method: | Quantitative |
| Detection Method: | Chemiluminescent |
| Sensitivity: | < 9.26 pg/mL |

Target Details

| | |
|-----------|----------------------------------|
| Target: | Relaxin |
| Abstract: | Relaxin Products |

Application Details

| | |
|--------------------|---|
| Application Notes: | <p>Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.</p> <p>Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.</p> <p>Standard Form: Lyophilized</p> |
|--------------------|---|

| | |
|--------|------------|
| Plate: | Pre-coated |
|--------|------------|

| | |
|---------------|-----------------------|
| Restrictions: | For Research Use only |
|---------------|-----------------------|

Handling

| | |
|----------|-------------|
| Storage: | 4 °C/-20 °C |
|----------|-------------|

| | |
|------------------|---|
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. |
|------------------|---|

| | |
|--------------|----------|
| Expiry Date: | 6 months |
|--------------|----------|