

Datasheet for ABIN6202997

VEGF183 CLIA Kit



| _ | | | | | |
|---|---|---|----|----|---|
| | W | 0 | rv | 10 | W |

| Quantity: | 96 tests | |
|--------------------------|---|--|
| Target: | VEGF183 | |
| Reactivity: | Human | |
| Method Type: | Sandwich ELISA | |
| Detection Range: | 15.62 pg/mL - 1000 pg/mL | |
| Minimum Detection Limit: | 15.62 pg/mL | |
| Application: | ELISA | |
| Product Details | | |
| Purpose: | Human Vascular Endothelial Growth Factor 183 (VEGF183) CLIA Kit is a Sandwich CLIA Kit for | |
| | use with Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other | |
| | biological fluids. | |
| Sample Type: | Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate | |
| Analytical Method: | Quantitative | |
| Detection Method: | Chemiluminescent | |
| Sensitivity: | < 5.8 pg/mL | |
| Target Details | | |
| Target: | VEGF183 | |
| Alternative Name: | Vascular Endothelial Growth Factor 183 | |

Application Details

| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less | | |
|--------------------|---|--|--|
| | than 5 % within the expiration date under appropriate storage conditions. To minimize | | |
| | performance fluctuations, operation procedures and lab conditions should be strictly controlled. | | |
| | It is also strongly suggested that the whole assay is performed by the same user throughout. | | |
| | Recommended dilutions: Optimal dilutions/concentrations should be determined by the end | | |
| | user. | | |
| | Standard Form: Lyophilized | | |
| Plate: | Pre-coated | | |
| Restrictions: | For Research Use only | | |
| | | | |
| Handling | | | |
| Storage: | 4 °C/-20 °C | | |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. | | |
| Expiry Date: | 6 months | | |