

Datasheet for ABIN6206331 NUP62 CLIA Kit



Overview

| Quantity: | 96 tests |
|--------------------------|--|
| Target: | NUP62 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.312 ng/mL - 20 ng/mL |
| Minimum Detection Limit: | 0.312 ng/mL |
| Application: | ELISA |
| Product Details | |
| Purpose: | Human Nucleoporin 62 (NUP62) CLIA Kit is a Sandwich CLIA Kit for use with Tissue |
| | homogenates and other biological fluids. |
| Sample Type: | Tissue Homogenate |
| Analytical Method: | Quantitative |

| Analytical Method: | Quantitative |
|--------------------|------------------|
| Detection Method: | Chemiluminescent |
| Sensitivity: | < 0.112 ng/mL |

Target Details

| Target: | NUP62 |
|-------------------|--|
| Alternative Name: | Nucleoporin 62kDa (NUP62 Products) |
| Pathways: | EGFR Signaling Pathway, SARS-CoV-2 Protein Interactome |

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/2 | Product datasheet for ABIN6206331 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

| Application Details | |
|---------------------|--|
| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user. Standard Form: Lyophilized |
| Plate: | Pre-coated |
| Restrictions: | For Research Use only |
| Handling | |
| Storage: | 4 °C/-20 °C |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. |
| Expiry Date: | 6 months |