

## Datasheet for ABIN6206397

## **NPTXR CLIA Kit**



Overview

Overview	
Quantity:	96 tests
Target:	NPTXR
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA
Product Details	
Purpose:	Human Neuronal Pentraxin Receptor (NPTXR) CLIA Kit is a Sandwich CLIA Kit for use with Tissue homogenates and other biological fluids.
Purpose:  Sample Type:	
	Tissue homogenates and other biological fluids.
Sample Type:	Tissue homogenates and other biological fluids.  Tissue Homogenate
Sample Type: Analytical Method:	Tissue homogenates and other biological fluids.  Tissue Homogenate  Quantitative
Sample Type:  Analytical Method:  Detection Method:	Tissue homogenates and other biological fluids.  Tissue Homogenate  Quantitative  Chemiluminescent
Sample Type:  Analytical Method:  Detection Method:  Sensitivity:	Tissue homogenates and other biological fluids.  Tissue Homogenate  Quantitative  Chemiluminescent

## **Application Details**

Application Notes:	Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less
	than 5 % within the expiration date under appropriate storage conditions. To minimize
	performance fluctuations, operation procedures and lab conditions should be strictly controlled.
	It is also strongly suggested that the whole assay is performed by the same user throughout.
	Recommended dilutions: Optimal dilutions/concentrations should be determined by the end
	user.
	Standard Form: Lyophilized
Plate:	Pre-coated
Restrictions:	For Research Use only
Handling	
Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months