

## Datasheet for ABIN6208121 **TXNRD2 CLIA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	TXNRD2
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

### Product Details

Purpose:	Human Thioredoxin Reductase 2 (TrxR2) CLIA Kit is a Sandwich CLIA Kit for use with Tissue homogenates, cell lysates and other biological fluids.
Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Sensitivity:	< 0.124 ng/mL

### Target Details

Target:	TXNRD2
Alternative Name:	Thioredoxin Reductase 2 ( <a href="#">TXNRD2 Products</a> )
Pathways:	<a href="#">Cell RedoxHomeostasis</a>

## Application Details

---

**Application Notes:**                      **Stability:** The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.

**Recommended dilutions:** Optimal dilutions/concentrations should be determined by the end user.

**Standard Form:** Lyophilized

---

**Plate:**    Pre-coated

---

**Restrictions:**                                      For Research Use only

---

## Handling

---

**Storage:**    4 °C/-20 °C

---

**Storage Comment:**                                      Upon receipt, store the kit according to the storage instruction in the kit's manual.

---

**Expiry Date:**    6 months

---