

Datasheet for ABIN6208346
Uroplakin 3A CLIA Kit



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	Uroplakin 3A (UPK3A)
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	15.62 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.62 pg/mL
Application:	ELISA

Product Details

Purpose:	Rat Uroplakin 3A (UPK3A) CLIA Kit is a Sandwich CLIA Kit for use with Serum, plasma, tissue homogenates and other biological fluids.
Sample Type:	Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Sensitivity:	< 6.5 pg/mL

Target Details

Target:	Uroplakin 3A (UPK3A)
Alternative Name:	Uroplakin 3A (UPK3A Products)

Application Details

Application Notes: **Stability:** The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Standard Form: Lyophilized

Plate: Pre-coated

Restrictions: For Research Use only

Handling

Storage: 4 °C/-20 °C

Storage Comment: Upon receipt, store the kit according to the storage instruction in the kit's manual.

Expiry Date: 6 months
