

## Datasheet for ABIN6209426

# **ADRA2C CLIA Kit**



_						
	V	$\triangle$	r۱	/1	$\triangle$	Λ/
	' V '		ΙV			v v

Quantity:	96 tests
Target:	ADRA2C
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.625 ng/mL - 40 ng/mL
Minimum Detection Limit:	0.625 ng/mL
Application:	ELISA

### **Product Details**

Purpose:	Human Adrenergic Receptor alpha 2C (ADRa2C) CLIA Kit is a Sandwich CLIA Kit for use with Tissue homogenates and other biological fluids.
Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Sensitivity:	< 0.237 ng/mL

### **Target Details**

Target:	ADRA2C
Alternative Name:	Adrenergic Receptor alpha 2C (ADRa2C) (ADRA2C Products)
Pathways <sup>.</sup>	FGER Signaling Pathway cAMP Metabolic Process

## **Application Details**

Application Notes:	Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less		
	than 5 % within the expiration date under appropriate storage conditions. To minimize		
	performance fluctuations, operation procedures and lab conditions should be strictly controlled.		
	It is also strongly suggested that the whole assay is performed by the same user throughout.		
	Recommended dilutions: Optimal dilutions/concentrations should be determined by the end		
	user.		
	Standard Form: Lyophilized		
Plate:	Pre-coated		
Restrictions:	For Research Use only		
Handling			
Storage:	4 °C/-20 °C		
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.		
Expiry Date:	6 months		