

Datasheet for ABIN6210159

RNF39 CLIA Kit



| () | ve | rvi | 6 | W |
|--------|-----|-------|--------|-----|
| \sim | v C | 1 V I | \sim | v v |

| Quantity: | 96 tests |
|--------------------------|--|
| Target: | RNF39 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.312 ng/mL - 20 ng/mL |
| Minimum Detection Limit: | 0.312 ng/mL |
| Application: | ELISA |
| Product Details | |
| Purpose: | Human Ring Finger Protein 39 (RNF39) CLIA Kit is a Sandwich CLIA Kit for use with Tissue |
| | homogenates and other biological fluids. |
| Sample Type: | Tissue Homogenate |
| Analytical Method: | Quantitative |
| Detection Method: | Chemiluminescent |
| Sensitivity: | < 0.125 ng/mL |
| Target Details | |
| Target: | RNF39 |
| Abstract: | RNF39 Products |

Application Details

| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less | | |
|--------------------|---|--|--|
| | than 5 % within the expiration date under appropriate storage conditions. To minimize | | |
| | performance fluctuations, operation procedures and lab conditions should be strictly controlled. | | |
| | It is also strongly suggested that the whole assay is performed by the same user throughout. | | |
| | Recommended dilutions: Optimal dilutions/concentrations should be determined by the end | | |
| | user. | | |
| | Standard Form: Lyophilized | | |
| Plate: | Pre-coated | | |
| Restrictions: | For Research Use only | | |
| | | | |
| Handling | | | |
| Storage: | 4 °C/-20 °C | | |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. | | |
| Expiry Date: | 6 months | | |