

## Datasheet for ABIN6222749 **CYP2C18 CLIA Kit**



[Go to Product page](#)

### Overview

|                          |                        |
|--------------------------|------------------------|
| Quantity:                | 96 tests               |
| Target:                  | CYP2C18                |
| Reactivity:              | Human                  |
| Method Type:             | Sandwich ELISA         |
| Detection Range:         | 0.625 ng/mL - 40 ng/mL |
| Minimum Detection Limit: | 0.625 ng/mL            |
| Application:             | ELISA                  |

### Product Details

|                    |   |
|--------------------|---|
| Purpose:           | Cytochrome P450 2C18 CLIA Kit is a Sandwich CLIA Kit for use with Tissue homogenates, cell lysates and other biological fluids. |
| Sample Type:       | Cell Lysate, Tissue Homogenate  |
| Analytical Method: | Quantitative  |
| Detection Method:  | Chemiluminescent  |
| Sensitivity:       | < 0.264 ng/mL   |

### Target Details

|                   |   |
|-------------------|---|
| Target:           | CYP2C18   |
| Alternative Name: | Cytochrome P450 2C18 ( <a href="#">CYP2C18 Products</a> ) |

## Application Details

---

**Application Notes:**                      **Stability:** The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.

**Recommended dilutions:** Optimal dilutions/concentrations should be determined by the end user.

**Standard Form:** Lyophilized

---

**Plate:**    Pre-coated

---

**Restrictions:**                                      For Research Use only

---

## Handling

---

**Storage:**    4 °C/-20 °C

---

**Storage Comment:**                                      Upon receipt, store the kit according to the storage instruction in the kit's manual.

---

**Expiry Date:**    6 months

---