

Datasheet for ABIN6224196

OBP2A CLIA Kit



Overview

Overview	
Quantity:	96 tests
Target:	OBP2A
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.781 ng/mL - 50 ng/mL
Minimum Detection Limit:	0.781 ng/mL
Application:	ELISA
Product Details	
Purpose:	Human Odorant Binding Protein 2A (OBP2A) CLIA Kit is a Sandwich CLIA Kit for use with Saliva,
	tissue homogenates and other biological fluids.
Sample Type:	
Sample Type: Analytical Method:	tissue homogenates and other biological fluids.
	tissue homogenates and other biological fluids. Saliva, Tissue Homogenate
Analytical Method:	tissue homogenates and other biological fluids. Saliva, Tissue Homogenate Quantitative
Analytical Method: Detection Method:	tissue homogenates and other biological fluids. Saliva, Tissue Homogenate Quantitative Chemiluminescent
Analytical Method: Detection Method: Sensitivity:	tissue homogenates and other biological fluids. Saliva, Tissue Homogenate Quantitative Chemiluminescent

Application Details

Application Notes:	Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less
	than 5 % within the expiration date under appropriate storage conditions. To minimize
	performance fluctuations, operation procedures and lab conditions should be strictly controlled.
	It is also strongly suggested that the whole assay is performed by the same user throughout.
	Recommended dilutions: Optimal dilutions/concentrations should be determined by the end
	user.
	Standard Form: Lyophilized
Plate:	Pre-coated
Restrictions:	For Research Use only
Handling	
Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months