

Datasheet for ABIN6226102

ADAM22 CLIA Kit



Overview

| Quantity: | 96 tests |
|--------------------------|------------------------|
| Target: | ADAM22 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.156 ng/mL - 10 ng/mL |
| Minimum Detection Limit: | 0.156 ng/mL |
| Application: | ELISA |
| | |

Product Details

| Purpose: | Human ADAM22 CLIA Kit is a Sandwich CLIA Kit for use with Serum, plasma, tissue homogenates and other biological fluids. |
|--------------------|--|
| Sample Type: | Plasma, Serum, Tissue Homogenate |
| Analytical Method: | Quantitative |
| Detection Method: | Chemiluminescent |
| Sensitivity: | < 0.058 ng/mL |

Target Details

| Target: | ADAM22 |
|-------------------|--------------------------|
| Alternative Name: | ADAM22 (ADAM22 Products) |
| Gene ID: | 53 |

Target Details UniProt: Q9P0K1 **Application Details** Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less Application Notes: than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user. Standard Form: Lyophilized Plate: Pre-coated Restrictions: For Research Use only

| Handling | |
|------------------|---|
| Storage: | 4 °C/-20 °C |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. |

6 months

Expiry Date: