

Datasheet for ABIN6229498

OGN CLIA Kit



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	OGN
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	31.25 pg/mL - 2000 pg/mL
Minimum Detection Limit:	31.25 pg/mL
Application:	ELISA

Product Details

Purpose:	Human Osteoglycin (OGN) CLIA Kit is a Sandwich CLIA Kit for use with Serum, plasma, tissue homogenates, cell lysates and other biological fluids.
Sample Type:	Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Sensitivity:	< 12.5 pg/mL

Target Details

Target:	OGN
Alternative Name:	Osteoglycin (OGN Products)
UniProt:	P20774 , A8K0R3 , Q7Z532

Target Details

Pathways: [Glycosaminoglycan Metabolic Process](#)

Application Details

Application Notes: Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Standard Form: Lyophilized

Plate: Pre-coated

Restrictions: For Research Use only

Handling

Storage: 4 °C/-20 °C

Storage Comment: Upon receipt, store the kit according to the storage instruction in the kit's manual.

Expiry Date: 6 months